


Work Package 1

PTB Highlights

Alexandra Bourguin, Ralf-Peter Kapsch

Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Lower Saxony, DE



SI-traceable primary standards for absorbed dose measurements, towards the development of a primary standard

- Development, optimisation and commissioning of reference radiation fields for electron beams with ultra-high dose per pulse

SI-traceable primary standards for absorbed dose measurements, towards the development of a primary standard

- Development, optimisation and commissioning of reference radiation fields for electron beams with ultra-high dose per pulse

- Development and testing of primary standards for absorbed dose in electron beams with ultra-high dose per pulse

- The PTB's initial beam characteristic:
 - Primary electron beam energy: 20 MeV
 - Pulse rate frequency: 5 Hz
 - Pulse length: 2.5 μ s

- The PTB's initial beam characteristic:
 - Primary electron beam energy: 20 MeV
 - Pulse rate frequency: 5 Hz
 - Pulse length: 2.5 μ s
 - Scattering: 100 μ m Cu exit window and a “scintillating” window
 - SSD of 70 cm
 - Size (Gaussian shape): 115 mm
 - Dose per pulse range: 0.4 Gy per pulse to 3.2 Gy per pulse

- The PTB's initial beam characteristic:
 - Primary electron beam energy: 20 MeV
 - Pulse rate frequency: 5 Hz
 - Pulse length: 2.5 μ s
 - Scattering: 100 μ m Cu exit window and a “scintillating” window
 - SSD of 70 cm
 - Size (Gaussian shape): 115 mm
 - Dose per pulse range: 0.4 Gy per pulse to 3.2 Gy per pulse
 - No Monte Carlo beam model

Optimisation of the exiting beam:

- 1- Characterize the electron beam in the vacuum beamline of the research linac.
Baseline to construct the Monte Carlo beam model

Optimisation of the exiting beam:

1- Characterize the electron beam in the vacuum beamline of the research linac.

2- Create an initial reference beam to validate the Monte Carlo beam model

By removing the scintillating window, able to reach a ~80 mm beam size (still Gaussian) with a dose rate between 1 Gy to 6 Gy per pulse.

Optimisation of the exiting beam:

1- Characterize the electron beam in the vacuum beamline of the research linac.

2- Create an initial reference beam to validate the Monte Carlo beam model

By removing the scintillating window, able to reach a ~80 mm beam size (still Gaussian) with a dose rate between 1 Gy to 6 Gy per pulse.

3- Validate the Monte Carlo beam model by comparing depth dose curve and profile

0.5 mm agreement on the R₅₀ depth evaluated and 2 mm agreement on the beam size

Optimisation of the exiting beam:

- 1- Characterize the electron beam in the vacuum beamline of the research linac.
- 2- Create an initial reference beam to validate the Monte Carlo beam model
By removing the scintillating window, able to reach a ~80 mm beam size (still Gaussian) with a dose rate between 1 Gy to 6 Gy per pulse.
- 3- Validate the Monte Carlo beam model by comparing depth dose curve and profile
0.5 mm agreement on the R₅₀ depth evaluated and 2 mm agreement on the beam size
- 4- Using the Monte Carlo beam model, 6 different beam set-ups were created and measured.
Enabling dose rate range between 0.1 Gy to 15 Gy per pulse

Optimisation of the exiting beam:

- 1- Characterize the electron beam in the vacuum beamline of the research linac.
- 2- Create an initial reference beam to validate the Monte Carlo beam model
By removing the scintillating window, able to reach a ~80 mm beam size (still Gaussian) with a dose rate between 1 Gy to 6 Gy per pulse.
- 3- Validate the Monte Carlo beam model by comparing depth dose curve and profile
0.5 mm agreement on the R₅₀ depth evaluated and 2 mm agreement on the beam size
- 4- Using the Monte Carlo beam model, 6 different beam set-up were created and measured.
Enabling dose rate range between 0.1 Gy to 15 Gy per pulse

-> Publication in PMB journal in 2022

Reference UHPDR electron beams:

- Two beams were established as reference;
 - DPP: 0.1 Gy to 6 Gy per pulse (5 Hz, 2.5 μ s)
 - Countless relative beam measurements with ion chambers, diamond detectors, diode
 - Six absolute dose measurement characterisations with alanine and two using water calorimeter

Reference UHPDR electron beams:

- Two beams were established as reference;
DPP: 0.1 Gy to 6 Gy per pulse (5 Hz, 2.5 μ s)
 - Countless relative beam measurements with ion chamber, diamond detector, diode detector
 - Six absolute dose measurement characterisations with alanine and two using water calorimeter

- A third beam setup, DPP up to 15 Gy (5 Hz, 2.5 μ s)
 - Large number of relative measurements with diamond detector and TimePix3 (out-of-field)
 - Three absolute dose measurement characterisations with alanine pellets

Reference UHPDR electron beams:

- Two beams were established as reference;
DPP: 0.1 Gy to 6 Gy per pulse (5 Hz, 2.5 μ s)
 - Countless relative beam measurement with ion chamber, diamond detector, diode detector
 - Six absolute dose measurement characterisation with Alanine and two using Water calorimeter

- A third beam setup, DPP up to 15 Gy (5 Hz, 2.5 μ s)
 - Large number of relative measurements with diamond detector and TimePix3 (out-of-field)
 - Three absolute dose measurement characterisations with alanine pellets

-> Pulse length can be modulated between about 1 μ s to 3 μ s

- Develop and test primary standards for absorbed dose in electron beams with ultra-high dose per pulse:

PTB's primary standard is the water calorimeter

$$D_w = \Delta T \cdot c_w \cdot K_{HD} \cdot K_{HC} \cdot K_p \cdot K_{rp} \cdot K_{T,depth} \cdot K_e,$$


Parameter		
ΔT	Radiation-induced temperature rise	Validated the measurement procedure
c_w	Specific heat capacity of water at 4 °C	4206.8(1.3) J kg ⁻¹ K ⁻¹
K_{HD}	Heat defect correction factor	Independent of the dose rate; 1.0000(14)
K_{HC}	Heat transfer correction factor	To be determined by FEM simulations
K_p	Field perturbation correction factor	MC simulations
K_{rp}	Radial beam profile correction factor	MC simulations
$K_{T,depth}$	Water temperature and depth correction factor	Calculation and MC simulations
K_e	Self-heating correction factor	Independent of the dose rate; measured

$$D_w = \Delta T \cdot c_w \cdot K_{HD} \cdot K_{HC} \cdot K_p \cdot K_{rp} \cdot K_{T,depth} \cdot K_e,$$

Parameter		
ΔT	Radiation-induced temperature rise	Validated the measurement procedure
c_w	Specific heat capacity of water at 4 °C	4206.8(1.3) J kg ⁻¹ K ⁻¹
K_{HD}	Heat defect correction factor	Independent of the dose rate; 1.0000(14)
K_{HC}	Heat transfer correction factor	To be determined by FEM simulations
K_p	Field perturbation correction factor	MC simulations
K_{rp}	Radial beam profile correction factor	MC simulations
$K_{T,depth}$	Water temperature and depth correction factor	Calculation and MC simulations
K_e	Self-heating correction factor	Independent of the dose rate; measured

Development of primary standards for absorbed dose in electron beams

1- Perform thermal simulations and measurements to validate them along with the protocol



UNPUBLISHED DATA-DO NOT COPY OR DISTRIBUTE

2- Calculate the correction factors required to perform absolute dosimetry in UHDR

Parameter	Conventional electron beam (20 MeV)		UHPDR electron beam (20 MeV)	
	Value	Uncertainty, k=1 [%]	Value	Uncertainty, k=1 [%]
ΔT [mK]	0.95	0.12	1.32 - 2.26	0.22
c_w [$J \cdot kg^{-1} \cdot K^{-1}$]	4206.8	0.03	4206.8	0.03
K_{HD}	1.0000	0.14	1.0000	0.14
K_{HC}	0.9965 – 1.0027	0.23	0.9965 - 1.0040	0.20
K_p	1.0005	0.15	0.9982	0.25
K_{rp}	1.0007	0.05	0.9982 – 1.0000	0.25
$K_{T,depth}$	0.9996 - 1.0004	0.02	1.0006 – 1.0011	0.05
K_e	1.0004	0.05	1.0000	0.05
Combined Correction factors	0.9977 – 1.0047	0.34	0.9953 – 1.0015	0.49

2- Calculate the correction factors required to perform absolute dosimetry in UHDR

Parameter	Conventional electron beam (20 MeV)		UHPDR electron beam (20 MeV)	
	Value	Uncertainty, k=1 [%]	Value	Uncertainty, k=1 [%]
ΔT [mK]	0.95	0.12	1.32 - 2.26	0.22
c_w [J·kg ⁻¹ ·K ⁻¹]	4206.8	0.03	4206.8	0.03
K_{HD}	1.0000	0.14	1.0000	0.14
K_{HC}	0.9965 – 1.0027	0.23	0.9965 - 1.0040	0.20
K_p	1.0005	0.15	0.9982	0.25
K_{rp}	1.0007	0.05	0.9982 – 1.0000	0.25
$K_{T,depth}$	0.9996 - 1.0004	0.02	1.0006 – 1.0011	0.05
K_e	1.0004	0.05	1.0000	0.05
Combined Correction factors	0.9977 – 1.0047	0.34	0.9953 – 1.0015	0.49

2- Calculate the correction factors required to perform absolute dosimetry in UHDR

Parameter	Conventional electron beam (20 MeV)		UHPDR electron beam (20 MeV)	
	Value	Uncertainty, k=1 [%]	Value	Uncertainty, k=1 [%]
ΔT [mK]	0.95	0.12	1.32 - 2.26	0.22
c_w [J·kg ⁻¹ ·K ⁻¹]	4206.8	0.03	4206.8	0.03
K_{HD}	1.0000	0.14	1.0000	0.14
K_{HC}	0.9965 – 1.0027	0.23	0.9965 - 1.0040	0.20
K_p	1.0005	0.15	0.9982	0.25
K_{rp}	1.0007	0.05	0.9982 – 1.0000	0.25
$K_{T,depth}$	0.9996 - 1.0004	0.02	1.0006 – 1.0011	0.05
K_e	1.0004	0.05	1.0000	0.05
Combined Correction factors	0.9977 – 1.0047	0.34	0.9953 – 1.0015	0.49

2- Calculate the correction factors required to perform absolute dosimetry in UHDR

Parameter	Conventional electron beam (20 MeV)		UHPDR electron beam (20 MeV)	
	Value	Uncertainty, k=1 [%]	Value	Uncertainty, k=1 [%]
ΔT [mK]	0.95	0.12	1.32 - 2.26	0.22
c_w [J·kg ⁻¹ ·K ⁻¹]	4206.8	0.03	4206.8	0.03
K_{HD}	1.0000	0.14	1.0000	0.14
K_{HC}	0.9965 – 1.0027	0.23	0.9965 - 1.0040	0.20
K_p	1.0005	0.15	0.9982	0.25
K_{rp}	1.0007	0.05	0.9982 – 1.0000	0.25
$K_{T,depth}$	0.9996 - 1.0004	0.02	1.0006 – 1.0011	0.05
K_e	1.0004	0.05	1.0000	0.05
Combined Correction factors	0.9977 – 1.0047	0.34	0.9953 – 1.0015	0.49

2- Calculate the correction factors required to perform absolute dosimetry in UHDR

Parameter	Conventional electron beam (20 MeV)		UHPDR electron beam (20 MeV)	
	Value	Uncertainty, k=1 [%]	Value	Uncertainty, k=1 [%]
ΔT [mK]	0.95	0.12	1.32 - 2.26	0.22
c_w [J·kg ⁻¹ ·K ⁻¹]	4206.8	0.03	4206.8	0.03
K_{HD}	1.0000	0.14	1.0000	0.14
K_{HC}	0.9965 – 1.0027	0.23	0.9965 - 1.0040	0.20
K_p	1.0005	0.15	0.9982	0.25
K_{rp}	1.0007	0.05	0.9982 – 1.0000	0.25
$K_{T,depth}$	0.9996 - 1.0004	0.02	1.0006 – 1.0011	0.05
K_e	1.0004	0.05	1.0000	0.05
Combined Correction factors	0.9977 – 1.0047	0.34	0.9953 – 1.0015	0.49

2- Calculate the correction factors required to perform absolute dosimetry in UHDR

Parameter	Conventional electron beam (20 MeV)		UHPDR electron beam (20 MeV)	
	Value	Uncertainty, k=1 [%]	Value	Uncertainty, k=1 [%]
ΔT [mK]	0.95	0.12	1.32 - 2.26	0.22
c_w [J·kg ⁻¹ ·K ⁻¹]	4206.8	0.03	4206.8	0.03
K_{HD}	1.0000	0.14	1.0000	0.14
K_{HC}	0.9965 – 1.0027	0.23	0.9965 - 1.0040	0.20
K_p	1.0005	0.15	0.9982	0.25
K_{rp}	1.0007	0.05	0.9982 – 1.0000	0.25
$K_{T,depth}$	0.9996 - 1.0004	0.02	1.0006 – 1.0011	0.05
K_e	1.0004	0.05	1.0000	0.05
Combined Correction factors	0.9977 – 1.0047	0.34	0.9953 – 1.0015	0.49

Development of primary standards for absorbed dose in electron beams



2- Calculate the correction factors required to perform absolute dosimetry in UHDR

-> Manuscript submitted to PMB journal in Dec. 2022

Parameter	Conventional electron beam (20 MeV)		UHPDR electron beam (20 MeV)	
	Value	Uncertainty, k=1 [%]	Value	Uncertainty, k=1 [%]
ΔT [mK]	0.95	0.12	1.32 - 2.26	0.22
c_w [J·kg ⁻¹ ·K ⁻¹]	4206.8	0.03	4206.8	0.03
K_{HD}	1.0000	0.14	1.0000	0.14
K_{HC}	0.9965 – 1.0027	0.23	0.9965 - 1.0040	0.20
K_p	1.0005	0.15	0.9982	0.25
K_{rp}	1.0007	0.05	0.9982 – 1.0000	0.25
$K_{T,depth}$	0.9996 - 1.0004	0.02	1.0006 – 1.0011	0.05
K_e	1.0004	0.05	1.0000	0.05
Combined Correction factors	0.9977 – 1.0047	0.34	0.9953 – 1.0015	0.49

UNPUBLISHED DATA-DO NOT COPY OR DISTRIBUTE

Validation of primary standards for absorbed dose in electron beams



- 3- Validate the absolute dose to water using the calorimeter by comparing with another primary standard




[Water calorimeter](#)

UNPUBLISHED DATA-DO NOT COPY OR DISTRIBUTE

Validation of primary standards for absorbed dose in electron beams

- 3- Validate the absolute dose to water using the calorimeter by comparing with another primary standard (Fricke from METAS, Deliverable 3)




Fricke dosimeter



Water calorimeter




Validation of primary standards for absorbed dose in electron beams



Fricke dosimeter





Fricke dosimeter



Big and heavy equipment
Requires new set of correction factors



Water calorimeter



Validation of primary standards for absorbed dose in electron beams



Water calorimeter

Fricke dosimeter





Travel with unstable chemical
or
Travel with a portable lab

Fricke dosimeter



Water calorimeter





Water calorimeter

Transfer dosimeter...

Easy to transport
Requires no additional expertise

Fricke dosimeter





Alanine dosimeter!

Fricke dosimeter




Water calorimeter





METAS

Validation of primary standards for absorbed dose in electron beams



Results



$$D_{\text{METAS/PTB}} = 1.002 \pm 0.012$$



Results



$$D_{\text{METAS/PTB}} = 1.002 \pm 0.012$$



$$D_{\text{PTB/NRC}} = 0.9997 \pm 0.0080$$

$$D_{\text{METAS/NRC}} = 1.002 \pm 0.011$$



- UHPDR electron reference beams were established, enabling dose rates between ~ 0.1 Gy per pulse up to ~ 15 Gy.

- UHPDR electron reference beams were established, enabling dose rates between ~ 0.1 Gy per pulse up to ~ 15 Gy.
- The PTB's water calorimeter has been established successfully with an uncertainty budget under initial expectation.

- UHPDR electron reference beams were established, enabling dose rates between ~ 0.1 Gy per pulse up to ~ 15 Gy.
- The PTB's water calorimeter has been established successfully with an uncertainty budget under initial expectation.
- PTB's and METAS's primary standard are consistent within stated uncertainty (1.002 ± 0.012).

All this with an average delay of just 6 months!
...much better than the Berlin airport construction delays



<http://uhdpulse-empir.eu/>

Acknowledgments

Franziska Frei	Malcolm McEwen
Thomas Hackel	Bryan Muir
Adrian Knyziak	Peter Peier
Rafael Kranzer	Andreas Schlesner
Achim Krauss	Andreas Schüller
Christoph Makowski	Olaf Tappe
Marco Marinelli	

This project 18HLT04 UHDpulse has received funding from the EMPIR programme co-financed by the Participating States and from the European Union's Horizon 2020 research and innovation programme.



The EMPIR initiative is co-funded by the European Union's Horizon 2020 research and innovation programme and the EMPIR Participating States



**Physikalisch-Technische Bundesanstalt
Braunschweig**

Bundesallee 100
38116 Braunschweig

Alexandra Bourgouin

Telefon: +49 531 592-6439

E-Mail: alexandra.bourgouin@ptb.de

www.ptb.de

Merci
Thank you
Dankeschön